

tion. The general assumption, that positive mesons always decay into positrons and neutrinos while negative mesons are mostly absorbed by nuclei, cannot be regarded as well founded.⁶ It seems to the writer that a coupling between a special counter arrangement similar to that designed by Rasetti⁷ and a cloud chamber will be suitable for this investigation.

¹ K. C. Wang, Phys. Rev. **61**, 97 (1942).

² J. S. Allen, Phys. Rev. **61**, 692 (1942).

³ E. J. Konopinski, Rev. Mod. Phys. **15**, 209 (1943).

⁴ G. C. Baldwin and G. S. Klaiber, Phys. Rev. **70**, 259 (1946).

⁵ H. R. Crane and J. Halpern, Phys. Rev. **56**, 232 (1939).

⁶ S. H. Neddermeyer and C. D. Anderson, Phys. Rev. **54**, 88 (1938);
Y. Nishina, M. Takeuchi, and T. Ichimiya, Phys. Rev. **55**, 585 (1939).

⁷ F. Rasetti, Phys. Rev. **60**, 198 (1941).

Velocity of Propagation of the Discharge in Geiger-Müller Counters

ERNEST WANTUCH

Department of Physics, New York University, New York, New York
March 24, 1947

AN important factor limiting the counting-rate readily achievable with counters is the time required for the avalanche to spread along the length of the central wire. When an ionizing event occurs in the volume of the counter, the electrons will be drawn toward the positive central wire. Near the central wire the electrons will acquire sufficient energy to ionize gas atoms by collision. During the chain of ionizations by collision, recombination, and excitation, photons, as well as new secondary electrons, are produced. The photons travel out in all directions, and eject electrons, thereby spreading the discharge down the length of the counter. The spread velocity of the discharge was measured by two experimental arrangements.

1. The first arrangement consisted of a counter with a divided cylinder. The main section of the copper counter cylinder was 30 cm long and 1 cm in diameter. At each end was placed another section of the same diameter, 1 cm long. The separation between the sections was less than 0.1 cm. One central wire served as anode for the three sections.

The outputs from the end sections were applied to a timing circuit.¹ The resulting pulse which represented the time interval between the discharges of the two end sections was applied to the vertical plates of a DuMont 208 oscilloscope. The background of the counter (approximately 200 counts per minute) was used as our source of impulses. This resulted in a random distribution of time intervals which were observed on the oscilloscope. The maximum time-duration resulted from a discharge initiated at one end of the counter and travelling the length of the counter to be recorded at the other end. The timing circuit was calibrated by a pulse generator giving a pair of pulses of known time separation.

A critical analysis of this experimental arrangement showed that the measurements represented the time taken by the discharge to spread along the length of the counter plus any fluctuations in the time required to collect the positive ions at the cylinder. For this reason, an arrangement designed to minimize these fluctuations was devised.

TABLE I. Velocity of propagation of the discharge.

Gas (cm of Hg)	Max. time interval	Spread velocity
	Method 1	
0.6 cm ethyl alcohol 9.4 cm argon	5.7 μ sec.	5.4 $\times 10^6$ cm/sec.
1.0 cm ethyl alcohol 9.4 cm argon	6.7	4.6 $\times 10^6$
2.0 cm ethyl alcohol 8.0 cm argon	8.0	3.9 $\times 10^6$
3.0 cm ethyl alcohol 7.0 cm argon	8.9	3.5 $\times 10^6$
	Method 2	
1.0 cm ethyl acetate 8.0 cm argon	4.5 μ sec.	6.2 $\times 10^6$ cm/sec.

2. A second counter was constructed consisting of a cylinder 30 cm long and 1 cm diameter. Small circular probes around the central wire were made of No. 26 platinum wire, and were of 0.14 cm inside diameter. The probes were located 28 cm apart.

Pulses obtained from these probes were separately amplified by one stage of 6SJ7 tubes, then combined in a suitable differentiating network. This resulted in a double pulse with the separation between the peaks representing the travel time of the discharge. The double pulse was further amplified by a video amplifier and applied to the vertical plates of a TS-28 synchroscope. The maximum time difference between the peaks can be interpreted in terms of the time scale of the sweeps. The experimental results are summarized in Table I. These results support the theory that the positive-ion sheath spreads by photon emission and ionization. When the proportion of organic vapor is increased, the travel time should increase as the organic molecule should absorb these photons and then pre-dissociate, thus removing the photon from the chain of ionizations. The fluctuations in the time required to collect positive ions at the cylinder must be comparatively small. Our results show a slower propagation velocity than results previously reported,² Huber, *et al.*, having found velocities approximately 50 percent higher than the values reported here for corresponding mixtures. One serious objection to their first method is the finite width of the pulses applied to the electrical shutters inside their counter. In their second method the sharp dependence of their experimental values on applied voltage cannot be accounted for.

The author wishes to thank Professor S. A. Korff for helpful suggestions.

¹ B. Rossi and N. Nereson, Rev. Sci. Inst. **17**, 65 (1946).

² P. Huber, F. Alden, and E. Baldinger, Helv. Phys. Acta **19**, 204 (1946).

Direction of Domain Magnetization in Powder Patterns

H. J. WILLIAMS

Bell Telephone Laboratories, Murray Hill, New Jersey

April 7, 1947

IN the course of an investigation of magnetic domain patterns, using colloidal particles of magnetite according to the well-known techniques,¹⁻⁴ a number of new patterns have been observed, and the directions of magnet-